

Haika NX435

The manufacturer: Globus [Shetland] Ltd.  
T2 Trafford Park, Twining Road,  
Trafford Park, MANCHESTER, M17 1SH. United Kingdom.

Confirms conformity to: PPE Regulation (EU) 2016/425  
And the standard(s): EN420:2003+A1:2009, EN374-1:2016, EN374-5:2016 & EN421:2010 (excluding clause 4.3)

Of the following PPE product(s):  
HAIKA NX435 – Five Fingered nitrile powder free disposable Gloves

The Notified Body SATRA Technology, [NB No. 2777]  
performed the EU Type-Examination, [Module B], and issued the EU Type-Examination Certificate:  
2777/10580-01/E02-01

The PPE is subject to the conformity assessment procedure, conformity to type based on internal production control plus supervised product checks at random intervals, [Module C2], under the surveillance of the Notified Body:  
SATRA Technology, [NB No. 2777]

Also

Medical Devices Directive 93/42/EEC as amended, Medical Devices Regulation (EU) 2017/745 as amended (effective from 26/05/2021), and with harmonised standards EN455-1:2000, EN455-2:2015, EN455-3:2015 and EN455-4:2009 and is Class I self-certified.

Signed for and on behalf of:  
Globus [Shetland] Ltd/, T2 Trafford Point, Twining Road,  
Trafford Park, MANCHESTER, M17 1SH. United Kingdom.



Name: Mr. Christian Halford  
Function: Regulatory & Quality Director

Date of issue: 07-Nov-22

EU Representative:  
Globus EMEA Ltd  
51 Dawson St  
Dublin  
D02 AN25 Ireland

HAIKA NX435

The manufacturer: Globus [Shetland] Ltd.  
T2 Trafford Park, Twining Road,  
Trafford Park, MANCHESTER, M17 1SH. United Kingdom.

Confirms conformity to: UK Regulation 2016/425 On PPE  
And the standard(s): EN420:2003+A1:2009, EN374-1:2016, EN374-5:2016 & EN421:2010 (excluding clause 4.3)

Of the following PPE product(s):  
HAIKA NX435 – Five Fingered nitrile powder Free disposable Gloves

Based on the Notified Body SATRA Technology, [NB No. 2777]  
performed the EU Type-Examination, [Module B], and issued the EU Type-Examination Certificate:  
2777/10580-01/E02-01

The PPE is subject to the conformity assessment procedure, conformity to type based on internal production control plus supervised product checks at random intervals, [Module C2], based on the surveillance from the Notified Body: SATRA Technology, [NB No. 2777]

Also

Medical Devices Directive 93/42/EEC as amended, Medical Devices Regulation (EU) 2017/745 as amended (effective from 26/05/2021), and with harmonised standards EN455-1:2000, EN455-2:2015, EN455-3:2015 and EN455-4:2009 and is Class I self-certified.

Signed for and on behalf of:  
Globus [Shetland] Ltd/, T2 Trafford Point, Twining Road,  
Trafford Park, MANCHESTER, M17 1SH. United Kingdom.



Name: Mr. Christian Halford  
Function: Regulatory & Quality Director

Date of issue: 07-Nov-22

D-o-C No.: HAIKA NX510\_NX520  
First Issued: 07/11/2022  
Issue No.: 01

UK/EU D-o-C (Internet Address): <https://gg-doc.com/Haika>