


EC Declaration of Conformity



We declare that the below mentioned product meets the provisions of EC Council Directive 93/42/EEC of 14th June 1993 concerning medical devices as amended by Directive 2007/47/EC.

All supporting documentation is retained under the premises of the manufacturer and notified body.

MANUFACTURER ADDRESS	Globus (Shetland) Ltd T2 Trafford Point Twining Road Trafford Park Manchester M17 1SH															
BRAND	 HAIKA															
PRODUCT	CALLISTO Surgical Gloves • Latex • Sterile • Powder Free															
	Product Codes: <table border="1"><tr><td>Size 5½</td><td>HAK2055</td><td>Size 7½</td><td>HAK2075</td></tr><tr><td>Size 6</td><td>HAK2060</td><td>Size 8</td><td>HAK2080</td></tr><tr><td>Size 6½</td><td>HAK2065</td><td>Size 8½</td><td>HAK2085</td></tr><tr><td>Size 7</td><td>HAK2070</td><td>Size 9</td><td>HAK2090</td></tr></table>	Size 5½	HAK2055	Size 7½	HAK2075	Size 6	HAK2060	Size 8	HAK2080	Size 6½	HAK2065	Size 8½	HAK2085	Size 7	HAK2070	Size 9
Size 5½	HAK2055	Size 7½	HAK2075													
Size 6	HAK2060	Size 8	HAK2080													
Size 6½	HAK2065	Size 8½	HAK2085													
Size 7	HAK2070	Size 9	HAK2090													
CLASSIFICATION (MDD 93/42, ANNEX IX)	Ila (Rule 7)															
CONFORMITY ASSESSMENT	Annex V, Annex VII															
STANDARDS	ISO 9001, ISO 13485, ISO 15223-1															
NOTIFIED BODY	BSI Kitemark Court Davy Avenue, Knowlhill Milton Keynes, MK5 8PP United Kingdom															
START OF CE-MARKING AND MDD COMPLIANCE	21/11/2018															

The product is intended to be used in surgical work and to be worn once and then discarded. Surgical glove is worn on the hand of surgeon and healthcare personnel to prevent contamination between healthcare personnel and the patient's body, fluids, waste or environment.

The gloves are designed for short-term use and are intended to be used in conjunction with invasive procedures.

Manchester, United Kingdom

Andrew Morris
Management Representative